



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC050123AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC063493.D

Aliphatic C9-C12	8038783.000	60.000	3.198	6.484	133979.717	137065.155	2.251
Aliphatic C12-C16	5341813.000	40.000	6.485	9.880	133545.325	135585.936	1.505
Aliphatic C16-C21	8195101.000	60.000	9.881	13.239	136585.017	135817.180	-0.565
Aliphatic C21-C28	10986248.000	80.000	13.240	16.891	137328.100	134266.069	-2.281
Aliphatic C28-C40	14950846.000	120.000	16.892	21.751	124590.383	131023.808	4.910
Aliphatic EPH	47512791.000	360.000	3.198	21.751	131979.975	134057.000	1.549