



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC050123AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC063507.D

Aliphatic C9-C12	9546646.000	60.000	3.198	6.484	159110.767	137065.155	-16.084
Aliphatic C12-C16	6429533.000	40.000	6.485	9.880	160738.325	135585.936	-18.551
Aliphatic C16-C21	9812128.000	60.000	9.881	13.239	163535.467	135817.180	-20.409
Aliphatic C21-C28	12817293.000	80.000	13.240	16.891	160216.163	134266.069	-19.327
Aliphatic C28-C40	17219965.000	120.000	16.892	21.751	143499.708	131023.808	-9.522
Aliphatic EPH	55825565.000	360.000	3.198	21.751	155071.014	134057.000	-15.675