



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC050123AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC063513.D

Aliphatic C9-C12	7596728.000	60.000	3.198	6.484	126612.133	137065.155	7.626
Aliphatic C12-C16	5063667.000	40.000	6.485	9.880	126591.675	135585.936	6.634
Aliphatic C16-C21	7787314.000	60.000	9.881	13.239	129788.567	135817.180	4.439
Aliphatic C21-C28	10450629.000	80.000	13.240	16.891	130632.863	134266.069	2.706
Aliphatic C28-C40	15634959.000	120.000	16.892	21.751	130291.325	131023.808	0.559
Aliphatic EPH	46533297.000	360.000	3.198	21.751	129259.158	134057.000	3.579