



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC050123AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC063517.D

Aliphatic C9-C12	7716150.000	60.000	3.198	6.484	128602.500	137065.155	6.174
Aliphatic C12-C16	5171050.000	40.000	6.485	9.880	129276.250	135585.936	4.654
Aliphatic C16-C21	7931918.000	60.000	9.881	13.239	132198.633	135817.180	2.664
Aliphatic C21-C28	10615287.000	80.000	13.240	16.891	132691.088	134266.069	1.173
Aliphatic C28-C40	15919924.000	120.000	16.892	21.751	132666.033	131023.808	-1.253
Aliphatic EPH	47354329.000	360.000	3.198	21.751	131539.803	134057.000	1.878