



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FE042723AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE041564.D

Aliphatic C9-C12	8739465.000	60.000	3.256	6.877	145657.750	149673.029	2.683
Aliphatic C12-C16	6131069.000	40.000	6.878	10.315	153276.725	155744.024	1.584
Aliphatic C16-C21	9224516.000	60.000	10.316	13.682	153741.933	161376.097	4.731
Aliphatic C21-C28	12258186.000	80.000	13.683	17.343	153227.325	158129.852	3.100
Aliphatic C28-C40	18131801.000	120.000	17.344	22.280	151098.342	139105.719	-8.621
Aliphatic EPH	54485037.000	360.000	3.256	22.280	151347.325	150654.953	-0.460