



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FE042723AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE041567.D

Aliphatic C9-C12	9629093.000	60.000	3.256	6.877	160484.883	149673.029	-7.224
Aliphatic C12-C16	6611659.000	40.000	6.878	10.315	165291.475	155744.024	-6.130
Aliphatic C16-C21	9798851.000	60.000	10.316	13.682	163314.183	161376.097	-1.201
Aliphatic C21-C28	12069902.000	80.000	13.683	17.343	150873.775	158129.852	4.589
Aliphatic C28-C40	16704069.000	120.000	17.344	22.280	139200.575	139105.719	-0.068
Aliphatic EPH	54813574.000	360.000	3.256	22.280	152259.928	150654.953	-1.065