



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FE042723AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE041589.D

Aliphatic C9-C12	8738901.000	60.000	3.256	6.877	145648.350	149673.029	2.689
Aliphatic C12-C16	6108028.000	40.000	6.878	10.315	152700.700	155744.024	1.954
Aliphatic C16-C21	9143190.000	60.000	10.316	13.682	152386.500	161376.097	5.571
Aliphatic C21-C28	12050098.000	80.000	13.683	17.343	150626.225	158129.852	4.745
Aliphatic C28-C40	18307200.000	120.000	17.344	22.280	152560.000	139105.719	-9.672
Aliphatic EPH	54347417.000	360.000	3.256	22.280	150965.047	150654.953	-0.206