



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FE042723AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE041620.D

Aliphatic C9-C12	9294319.000	60.000	3.256	6.877	154905.317	149673.029	-3.496
Aliphatic C12-C16	6429753.000	40.000	6.878	10.315	160743.825	155744.024	-3.210
Aliphatic C16-C21	9533958.000	60.000	10.316	13.682	158899.300	161376.097	1.535
Aliphatic C21-C28	11725012.000	80.000	13.683	17.343	146562.650	158129.852	7.315
Aliphatic C28-C40	16716996.000	120.000	17.344	22.280	139308.300	139105.719	-0.146
Aliphatic EPH	53700038.000	360.000	3.256	22.280	149166.772	150654.953	0.988