



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC080323AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC064306.D

Aliphatic C9-C12	11130020.000	60.000	3.205	6.477	185500.333	163010.873	-13.796
Aliphatic C12-C16	7126336.000	40.000	6.478	9.864	178158.400	157122.407	-13.388
Aliphatic C16-C21	10204070.000	60.000	9.865	13.215	170067.833	151210.353	-12.471
Aliphatic C21-C28	12536739.000	80.000	13.216	16.860	156709.238	138650.389	-13.025
Aliphatic C28-C40	17600197.000	120.000	16.861	21.693	146668.308	137334.939	-6.796
Aliphatic EPH	58597362.000	360.000	3.205	21.693	162770.450	146417.760	-11.169