



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC080323AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC064323.D

Aliphatic C9-C12	11499216.000	60.000	3.205	6.477	191653.600	163010.873	-17.571
Aliphatic C12-C16	7322714.000	40.000	6.478	9.864	183067.850	157122.407	-16.513
Aliphatic C16-C21	10358830.000	60.000	9.865	13.215	172647.167	151210.353	-14.177
Aliphatic C21-C28	12854726.000	80.000	13.216	16.860	160684.075	138650.389	-15.892
Aliphatic C28-C40	18587890.000	120.000	16.861	21.693	154899.083	137334.939	-12.789
Aliphatic EPH	60623376.000	360.000	3.205	21.693	168398.267	146417.760	-15.012