



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC080723AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC064340.D

Aliphatic C9-C12	11359572.000	60.000	3.207	6.479	189326.200	163010.873	-16.143
Aliphatic C12-C16	7131470.000	40.000	6.480	9.864	178286.750	157122.407	-13.470
Aliphatic C16-C21	10034939.000	60.000	9.865	13.215	167248.983	151210.353	-10.607
Aliphatic C21-C28	12391535.000	80.000	13.216	16.862	154894.188	138650.389	-11.716
Aliphatic C28-C40	16672979.000	120.000	16.863	21.694	138941.492	137334.939	-1.170
Aliphatic EPH	57590495.000	360.000	3.207	21.694	159973.597	146417.760	-9.258