



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FD080923AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD046310.D

Aliphatic C9-C12	8143979.000	60.000	3.125	6.380	135732.983	159963.790	15.148
Aliphatic C12-C16	5594893.000	40.000	6.381	9.755	139872.325	160907.029	13.073
Aliphatic C16-C21	8899825.000	60.000	9.756	13.095	148330.417	151641.397	2.183
Aliphatic C21-C28	12020213.000	80.000	13.096	16.729	150252.663	131757.393	-14.037
Aliphatic C28-C40	15645002.000	120.000	16.730	21.460	130375.017	118031.614	-10.458
Aliphatic EPH	50303912.000	360.000	3.125	21.460	139733.089	138436.049	-0.937