



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FD080923AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD046320.D

Aliphatic C9-C12	8014908.000	60.000	3.125	6.380	133581.800	159963.790	16.492
Aliphatic C12-C16	5505798.000	40.000	6.381	9.755	137644.950	160907.029	14.457
Aliphatic C16-C21	8723215.000	60.000	9.756	13.095	145386.917	151641.397	4.125
Aliphatic C21-C28	11803313.000	80.000	13.096	16.729	147541.413	131757.393	-11.980
Aliphatic C28-C40	15526579.000	120.000	16.730	21.460	129388.158	118031.614	-9.622
Aliphatic EPH	49573813.000	360.000	3.125	21.460	137705.036	138436.049	0.528