



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FE080723AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE043770.D

Aliphatic C9-C12	9408411.000	60.000	3.230	6.844	156806.850	164554.627	4.708
Aliphatic C12-C16	6535825.000	40.000	6.845	10.280	163395.625	172441.496	5.246
Aliphatic C16-C21	9903657.000	60.000	10.281	13.648	165060.950	178392.491	7.473
Aliphatic C21-C28	12337685.000	80.000	13.649	17.312	154221.063	172814.004	10.759
Aliphatic C28-C40	15792548.000	120.000	17.313	22.230	131604.567	145721.783	9.688
Aliphatic EPH	53978126.000	360.000	3.230	22.230	149939.239	163295.059	8.179