



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FE080723AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE043780.D

Aliphatic C9-C12	9275263.000	60.000	3.230	6.844	154587.717	164554.627	6.057
Aliphatic C12-C16	6442214.000	40.000	6.845	10.280	161055.350	172441.496	6.603
Aliphatic C16-C21	9806347.000	60.000	10.281	13.648	163439.117	178392.491	8.382
Aliphatic C21-C28	12330354.000	80.000	13.649	17.312	154129.425	172814.004	10.812
Aliphatic C28-C40	15716359.000	120.000	17.313	22.230	130969.658	145721.783	10.123
Aliphatic EPH	53570537.000	360.000	3.230	22.230	148807.047	163295.059	8.872