



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FE080723AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE043794.D

Aliphatic C9-C12	9465042.000	60.000	3.230	6.844	157750.700	164554.627	4.135
Aliphatic C12-C16	6576519.000	40.000	6.845	10.280	164412.975	172441.496	4.656
Aliphatic C16-C21	9857002.000	60.000	10.281	13.648	164283.367	178392.491	7.909
Aliphatic C21-C28	12103123.000	80.000	13.649	17.312	151289.038	172814.004	12.456
Aliphatic C28-C40	15646990.000	120.000	17.313	22.230	130391.583	145721.783	10.520
Aliphatic EPH	53648676.000	360.000	3.230	22.230	149024.100	163295.059	8.739