



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FE080723AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE043800.D

Aliphatic C9-C12	9539353.000	60.000	3.230	6.844	158989.217	164554.627	3.382
Aliphatic C12-C16	6636983.000	40.000	6.845	10.280	165924.575	172441.496	3.779
Aliphatic C16-C21	10025903.000	60.000	10.281	13.648	167098.383	178392.491	6.331
Aliphatic C21-C28	12369142.000	80.000	13.649	17.312	154614.275	172814.004	10.531
Aliphatic C28-C40	16003059.000	120.000	17.313	22.230	133358.825	145721.783	8.484
Aliphatic EPH	54574440.000	360.000	3.230	22.230	151595.667	163295.059	7.165