



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FF081023AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FF013178.D

|                   |              |         |        |        |            |            |       |
|-------------------|--------------|---------|--------|--------|------------|------------|-------|
| Aliphatic C9-C12  | 6218544.000  | 60.000  | 3.185  | 6.803  | 103642.400 | 107407.725 | 3.506 |
| Aliphatic C12-C16 | 4169363.000  | 40.000  | 6.804  | 10.235 | 104234.075 | 108171.933 | 3.640 |
| Aliphatic C16-C21 | 6341319.000  | 60.000  | 10.236 | 13.600 | 105688.650 | 109699.122 | 3.656 |
| Aliphatic C21-C28 | 8395514.000  | 80.000  | 13.601 | 17.257 | 104943.925 | 109229.526 | 3.923 |
| Aliphatic C28-C40 | 12062103.000 | 120.000 | 17.258 | 22.122 | 100517.525 | 105721.890 | 4.923 |
| Aliphatic EPH     | 37186843.000 | 360.000 | 3.185  | 22.122 | 103296.786 | 107717.436 | 4.104 |