



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FG081023AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG012666.D

Aliphatic C9-C12	6305861.000	60.000	3.246	6.869	105097.683	105192.935	0.091
Aliphatic C12-C16	4300905.000	40.000	6.870	10.315	107522.625	108120.753	0.553
Aliphatic C16-C21	6571098.000	60.000	10.316	13.690	109518.300	110346.515	0.751
Aliphatic C21-C28	8373701.000	80.000	13.691	17.358	104671.263	105912.016	1.171
Aliphatic C28-C40	10855911.000	120.000	17.359	22.311	90465.925	91132.051	0.731
Aliphatic EPH	36407476.000	360.000	3.246	22.311	101131.878	101850.012	0.705