



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FG081023AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG012676.D

Aliphatic C9-C12	6306434.000	60.000	3.246	6.869	105107.233	105192.935	0.081
Aliphatic C12-C16	4305045.000	40.000	6.870	10.315	107626.125	108120.753	0.457
Aliphatic C16-C21	6590885.000	60.000	10.316	13.690	109848.083	110346.515	0.452
Aliphatic C21-C28	8393385.000	80.000	13.691	17.358	104917.313	105912.016	0.939
Aliphatic C28-C40	10917993.000	120.000	17.359	22.311	90983.275	91132.051	0.163
Aliphatic EPH	36513742.000	360.000	3.246	22.311	101427.061	101850.012	0.415