



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

WATER GASOLINE RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLIC.

Lab Name: Chemtech Client: Chemtech Consulting Group  
Lab Code: CHEM Cas No: O3572 SAS No : O3572 SDG No: O3572  
Matrix Spike - EPA Sample No : BSF0711W2 Datafile: FB030160.D

COMPOUND	SPIKE ADDED ug/L	CONCENTRATION ug/L	LCS/LCSD CONCENTRATION ug/L	% REC	QC LIMITS
GRO	180	0	189	105	50-150

LCS/LCSD % Recovery RPD : 9.4