



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/04/2023 Initial Calibration Date(s): 07/20/2023 07/20/2023

Continuing Calib Time: 09:49 Initial Calibration Time(s): 15:20 16:57

GC Column: RTX-CLP ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|------------------|------------|-----------|------------------------|------|------------|
| DICAMBA          | 6.51       | 6.52      | 6.42                   | 6.62 | 0.01       |
| 2,4-DCAA         | 6.37       | 6.37      | 6.27                   | 6.47 | 0.00       |
| DICHLORPROP      | 7.10       | 7.11      | 7.01                   | 7.21 | 0.01       |
| 2,4-D            | 7.30       | 7.30      | 7.20                   | 7.40 | 0.00       |
| 2,4,5-TP(Silvex) | 8.05       | 8.06      | 7.96                   | 8.16 | 0.01       |
| 2,4,5-T          | 8.31       | 8.31      | 8.21                   | 8.41 | 0.00       |
| 2,4-DB           | 8.81       | 8.82      | 8.72                   | 8.92 | 0.01       |
| Dinoseb          | 9.87       | 9.87      | 9.77                   | 9.97 | 0.00       |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/04/2023 Initial Calibration Date(s): 07/20/2023 07/20/2023

Continuing Calib Time: 09:49 Initial Calibration Time(s): 15:20 16:57

GC Column: RTX-CLP2 ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|------------------|------------|-----------|------------------------|------|------------|
| DICAMBA          | 6.97       | 6.97      | 6.87                   | 7.07 | 0.00       |
| 2,4-DCAA         | 6.80       | 6.81      | 6.71                   | 6.91 | 0.01       |
| DICHLORPROP      | 7.58       | 7.59      | 7.49                   | 7.69 | 0.01       |
| 2,4-D            | 7.87       | 7.88      | 7.78                   | 7.98 | 0.01       |
| 2,4,5-TP(Silvex) | 8.66       | 8.66      | 8.56                   | 8.76 | 0.01       |
| 2,4,5-T          | 9.03       | 9.03      | 8.93                   | 9.13 | 0.00       |
| 2,4-DB           | 9.53       | 9.54      | 9.44                   | 9.64 | 0.01       |
| Dinoseb          | 9.86       | 9.87      | 9.77                   | 9.97 | 0.01       |