



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/04/2023 Initial Calibration Date(s): 07/20/2023 07/20/2023

Continuing Calib Time: 21:10 Initial Calibration Time(s): 15:20 16:57

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	6.51	6.52	6.42	6.62	0.01
2,4-DCAA	6.37	6.37	6.27	6.47	0.00
DICHLORPROP	7.10	7.11	7.01	7.21	0.01
2,4-D	7.30	7.30	7.20	7.40	0.00
2,4,5-TP(Silvex)	8.05	8.06	7.96	8.16	0.01
2,4,5-T	8.31	8.31	8.21	8.41	0.00
2,4-DB	8.81	8.82	8.72	8.92	0.01
Dinoseb	9.87	9.87	9.77	9.97	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/04/2023 Initial Calibration Date(s): 07/20/2023 07/20/2023

Continuing Calib Time: 21:10 Initial Calibration Time(s): 15:20 16:57

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	6.97	6.97	6.87	7.07	0.00
2,4-DCAA	6.80	6.81	6.71	6.91	0.01
DICHLORPROP	7.58	7.59	7.49	7.69	0.01
2,4-D	7.87	7.88	7.78	7.98	0.01
2,4,5-TP(Silvex)	8.66	8.66	8.56	8.76	0.01
2,4,5-T	9.03	9.03	8.93	9.13	0.00
2,4-DB	9.53	9.54	9.44	9.64	0.01
Dinoseb	9.86	9.87	9.77	9.97	0.01