



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/09/2023 Initial Calibration Date(s): 08/08/2023 08/08/2023

Continuing Calib Time: 14:07 Initial Calibration Time(s): 09:46 11:28

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	6.51	6.51	6.41	6.61	0.00
2,4-DCAA	6.37	6.37	6.27	6.47	0.00
DICHLORPROP	7.10	7.10	7.00	7.20	0.00
2,4-D	7.30	7.30	7.20	7.40	0.00
2,4,5-TP(Silvex)	8.05	8.05	7.95	8.15	0.00
2,4,5-T	8.30	8.30	8.20	8.40	0.00
2,4-DB	8.81	8.81	8.71	8.91	0.00
Dinoseb	9.86	9.87	9.77	9.97	0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/09/2023 Initial Calibration Date(s): 08/08/2023 08/08/2023

Continuing Calib Time: 14:07 Initial Calibration Time(s): 09:46 11:28

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	6.96	6.96	6.86	7.06	0.00
2,4-DCAA	6.80	6.80	6.70	6.90	0.00
DICHLORPROP	7.58	7.58	7.48	7.68	0.00
2,4-D	7.86	7.86	7.76	7.96	0.00
2,4,5-TP(Silvex)	8.65	8.65	8.55	8.75	0.00
2,4,5-T	9.02	9.02	8.92	9.12	0.00
2,4-DB	9.52	9.53	9.43	9.63	0.01
Dinoseb	9.85	9.85	9.75	9.95	0.00