



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 07/20/2023 07/20/2023

Client Sample No.: CCAL01 Date Analyzed: 08/04/2023

Lab Sample No.: HSTDCCC750 Data File : PS023951.D Time Analyzed: 21:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	8.306	8.210	8.410	752.810	712.500	5.7
2,4,5-TP(Silvex)	8.051	7.955	8.155	732.360	712.500	2.8
2,4-D	7.299	7.202	7.402	719.070	705.000	2.0
2,4-DB	8.812	8.718	8.918	723.710	712.500	1.6
2,4-DCAA	6.370	6.273	6.473	766.360	750.000	2.2
DICAMBA	6.513	6.415	6.615	733.990	705.000	4.1
DICHLORPROP	7.102	7.005	7.205	711.730	705.000	1.0
Dinoseb	9.867	9.774	9.974	791.040	705.000	12.2



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 07/20/2023 07/20/2023

Client Sample No.: CCAL01 Date Analyzed: 08/04/2023

Lab Sample No.: HSTDCCC750 Data File : PS023951.D Time Analyzed: 21:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.026	8.933	9.133	741.680	712.500	4.1
2,4,5-TP(Silvex)	8.655	8.562	8.762	744.110	712.500	4.4
2,4-D	7.869	7.775	7.975	683.930	705.000	-3.0
2,4-DB	9.530	9.437	9.637	767.190	712.500	7.7
2,4-DCAA	6.804	6.708	6.908	752.410	750.000	0.3
DICAMBA	6.967	6.872	7.072	736.590	705.000	4.5
DICHLORPROP	7.584	7.490	7.690	690.620	705.000	-2.0
Dinoseb	9.859	9.766	9.966	757.430	705.000	7.4