



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 07/20/2023 07/20/2023

Client Sample No.: CCAL02 Date Analyzed: 08/05/2023

Lab Sample No.: HSTDCCC750 Data File : PS023958.D Time Analyzed: 00:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	8.306	8.210	8.410	752.870	712.500	5.7
2,4,5-TP(Silvex)	8.051	7.955	8.155	740.350	712.500	3.9
2,4-D	7.298	7.202	7.402	719.820	705.000	2.1
2,4-DB	8.813	8.718	8.918	732.540	712.500	2.8
2,4-DCAA	6.370	6.273	6.473	772.560	750.000	3.0
DICAMBA	6.512	6.415	6.615	740.210	705.000	5.0
DICHLORPROP	7.102	7.005	7.205	719.230	705.000	2.0
Dinoseb	9.867	9.774	9.974	800.840	705.000	13.6



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 07/20/2023 07/20/2023

Client Sample No.: CCAL02 Date Analyzed: 08/05/2023

Lab Sample No.: HSTDCCC750 Data File : PS023958.D Time Analyzed: 00:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.025	8.933	9.133	746.450	712.500	4.8
2,4,5-TP(Silvex)	8.654	8.562	8.762	751.820	712.500	5.5
2,4-D	7.869	7.775	7.975	704.910	705.000	0.0
2,4-DB	9.530	9.437	9.637	772.050	712.500	8.4
2,4-DCAA	6.803	6.708	6.908	764.850	750.000	2.0
DICAMBA	6.967	6.872	7.072	745.550	705.000	5.8
DICHLORPROP	7.584	7.490	7.690	704.590	705.000	-0.1
Dinoseb	9.858	9.766	9.966	786.150	705.000	11.5