



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 08/08/2023 08/08/2023

Client Sample No.: CCAL03 Date Analyzed: 08/09/2023

Lab Sample No.: HSTDCCC750 Data File : PS023994.D Time Analyzed: 08:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	8.303	8.204	8.404	755.220	712.500	6.0
2,4,5-TP(Silvex)	8.048	7.948	8.148	734.400	712.500	3.1
2,4-D	7.296	7.196	7.396	725.080	705.000	2.8
2,4-DB	8.808	8.708	8.908	768.420	712.500	7.8
2,4-DCAA	6.367	6.268	6.468	757.230	750.000	1.0
DICAMBA	6.510	6.410	6.610	721.400	705.000	2.3
DICHLORPROP	7.099	6.999	7.199	712.210	705.000	1.0
Dinoseb	9.864	9.765	9.965	728.410	705.000	3.3



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 08/08/2023 08/08/2023

Client Sample No.: CCAL03 Date Analyzed: 08/09/2023

Lab Sample No.: HSTDCCC750 Data File : PS023994.D Time Analyzed: 08:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.019	8.920	9.120	746.810	712.500	4.8
2,4,5-TP(Silvex)	8.647	8.550	8.750	731.430	712.500	2.7
2,4-D	7.863	7.764	7.964	725.650	705.000	2.9
2,4-DB	9.522	9.425	9.625	742.960	712.500	4.3
2,4-DCAA	6.798	6.699	6.899	757.850	750.000	1.0
DICAMBA	6.962	6.863	7.063	722.890	705.000	2.5
DICHLORPROP	7.578	7.479	7.679	712.080	705.000	1.0
Dinoseb	9.851	9.754	9.954	714.830	705.000	1.4