



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 08/08/2023 08/08/2023

Client Sample No.: CCAL04 Date Analyzed: 08/09/2023

Lab Sample No.: HSTDCCC750 Data File : PS024003.D Time Analyzed: 14:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	8.303	8.204	8.404	758.650	712.500	6.5
2,4,5-TP(Silvex)	8.048	7.948	8.148	740.330	712.500	3.9
2,4-D	7.296	7.196	7.396	727.480	705.000	3.2
2,4-DB	8.808	8.708	8.908	761.730	712.500	6.9
2,4-DCAA	6.368	6.268	6.468	763.550	750.000	1.8
DICAMBA	6.510	6.410	6.610	728.330	705.000	3.3
DICHLORPROP	7.099	6.999	7.199	719.140	705.000	2.0
Dinoseb	9.864	9.765	9.965	728.270	705.000	3.3



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 08/08/2023 08/08/2023

Client Sample No.: CCAL04 Date Analyzed: 08/09/2023

Lab Sample No.: HSTDCCC750 Data File : PS024003.D Time Analyzed: 14:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.019	8.920	9.120	752.800	712.500	5.7
2,4,5-TP(Silvex)	8.648	8.550	8.750	742.300	712.500	4.2
2,4-D	7.863	7.764	7.964	733.050	705.000	4.0
2,4-DB	9.523	9.425	9.625	748.930	712.500	5.1
2,4-DCAA	6.799	6.699	6.899	770.330	750.000	2.7
DICAMBA	6.962	6.863	7.063	735.150	705.000	4.3
DICHLORPROP	7.578	7.479	7.679	723.440	705.000	2.6
Dinoseb	9.851	9.754	9.954	719.400	705.000	2.0