

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	08/05/23
Project:	LOD-LOQ Study	Date Received:	08/05/23
Client Sample ID:	PIBLK-PS023957.D	SDG No.:	O3572
Lab Sample ID:	I.BLK-PS023957.D	Matrix:	WATER
Analytical Method:	SW8151A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:		uL	
Extraction Type:		Test:	Herbicide
GPC Factor :	1.0	PH :	
Prep Method :	SW3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS023957.D	1		08/05/23	ps080423

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1918-00-9	DICAMBA	0.51	U	0.51	2.00	ug/L
120-36-5	DICHLORPROP	0.54	U	0.54	2.00	ug/L
94-75-7	2,4-D	0.57	U	0.57	2.00	ug/L
93-72-1	2,4,5-TP (Silvex)	0.54	U	0.54	2.00	ug/L
93-76-5	2,4,5-T	0.58	U	0.58	2.00	ug/L
94-82-6	2,4-DB	0.56	U	0.56	2.00	ug/L
88-85-7	DINOSEB	0.71	U	0.71	2.00	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	507		39 - 175	101%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit