



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Instrument ID: ECD_R Calibration Date(s): 08/08/2023 08/08/2023

Calibration Times: 16:30 19:12

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PR062493.D RT 750 = PR062494.D

RT 500 = PR062495.D RT 250 = PR062496.D RT 050 = PR062497.D

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Aroclor-1016-1	(1)	4.92	4.91	4.91	4.91	4.91	4.91	4.81	5.01
Aroclor-1016-2	(2)	4.94	4.94	4.94	4.94	4.94	4.94	4.84	5.04
Aroclor-1016-3	(3)	5.00	5.00	5.00	5.00	5.00	5.00	4.90	5.10
Aroclor-1016-4	(4)	5.10	5.10	5.10	5.10	5.10	5.10	5.00	5.20
Aroclor-1016-5	(5)	5.42	5.42	5.42	5.42	5.42	5.42	5.32	5.52
Aroclor-1260-1	(1)	6.64	6.64	6.64	6.64	6.64	6.64	6.54	6.74
Aroclor-1260-2	(2)	6.92	6.92	6.92	6.92	6.92	6.92	6.82	7.02
Aroclor-1260-3	(3)	7.31	7.31	7.31	7.31	7.31	7.31	7.21	7.41
Aroclor-1260-4	(4)	7.55	7.55	7.55	7.55	7.55	7.55	7.45	7.65
Aroclor-1260-5	(5)	7.89	7.89	7.89	7.89	7.89	7.89	7.79	7.99
Decachlorobiphenyl		9.62	9.62	9.62	9.62	9.62	9.62	9.52	9.72
Tetrachloro-m-xylene		3.69	3.68	3.69	3.69	3.68	3.69	3.59	3.79

RETENTION TIMES OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM

Case No.: 03572

SAS No.: 03572

SDG NO.: 03572

Instrument ID: ECD_R

Calibration Date(s): 08/08/2023 08/08/2023

Calibration Times: 16:30 19:12

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 =	<u>PR062493.D</u>	RT 750 =	<u>PR062494.D</u>
	RT 500 =	PR062495.D	RT 250 =	PR062496.D
			RT 050 =	PR062497.D

[illegible]