



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/08/2023 Initial Calibration Date(s): 08/07/2023 08/07/2023

Continuing Calib Time: 11:53 Initial Calibration Time(s): 16:22 23:59

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.60       | 5.59      | 5.49      | 5.69  | 0.00       |
| Aroclor-1016-2       | (2) | 5.62       | 5.61      | 5.51      | 5.71  | -0.01      |
| Aroclor-1016-3       | (3) | 5.68       | 5.68      | 5.58      | 5.78  | 0.00       |
| Aroclor-1016-4       | (4) | 5.78       | 5.78      | 5.68      | 5.88  | 0.00       |
| Aroclor-1016-5       | (5) | 6.08       | 6.08      | 5.98      | 6.18  | 0.00       |
| Aroclor-1260-1       | (1) | 7.24       | 7.23      | 7.13      | 7.33  | -0.01      |
| Aroclor-1260-2       | (2) | 7.50       | 7.49      | 7.39      | 7.59  | -0.01      |
| Aroclor-1260-3       | (3) | 7.87       | 7.86      | 7.76      | 7.96  | -0.01      |
| Aroclor-1260-4       | (4) | 8.10       | 8.10      | 8.00      | 8.20  | 0.00       |
| Aroclor-1260-5       | (5) | 8.44       | 8.43      | 8.33      | 8.53  | -0.01      |
| Tetrachloro-m-xylene |     | 4.41       | 4.41      | 4.31      | 4.51  | 0.00       |
| Decachlorobiphenyl   |     | 10.35      | 10.34     | 10.24     | 10.44 | -0.01      |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/08/2023 Initial Calibration Date(s): 08/07/2023 08/07/2023

Continuing Calib Time: 11:53 Initial Calibration Time(s): 16:22 23:59

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1016-1 (1)   | 4.71       | 4.71      | 4.61      | 4.81 | 0.00       |
| Aroclor-1016-2 (2)   | 4.73       | 4.73      | 4.63      | 4.83 | 0.00       |
| Aroclor-1016-3 (3)   | 4.90       | 4.90      | 4.80      | 5.00 | 0.00       |
| Aroclor-1016-4 (4)   | 4.95       | 4.95      | 4.85      | 5.05 | 0.00       |
| Aroclor-1016-5 (5)   | 5.16       | 5.16      | 5.06      | 5.26 | 0.00       |
| Aroclor-1260-1 (1)   | 6.20       | 6.20      | 6.10      | 6.30 | 0.00       |
| Aroclor-1260-2 (2)   | 6.38       | 6.39      | 6.29      | 6.49 | 0.01       |
| Aroclor-1260-3 (3)   | 6.54       | 6.54      | 6.44      | 6.64 | 0.00       |
| Aroclor-1260-4 (4)   | 7.01       | 7.01      | 6.91      | 7.11 | 0.00       |
| Aroclor-1260-5 (5)   | 7.26       | 7.26      | 7.16      | 7.36 | 0.00       |
| Tetrachloro-m-xylene | 3.62       | 3.62      | 3.52      | 3.72 | 0.00       |
| Decachlorobiphenyl   | 8.64       | 8.64      | 8.54      | 8.74 | 0.00       |