



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/08/2023 Initial Calibration Date(s): 08/07/2023 08/07/2023

Continuing Calib Time: 21:07 Initial Calibration Time(s): 16:22 23:59

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.60	5.59	5.49	5.69	-0.01
Aroclor-1242-2 (2)	5.62	5.61	5.51	5.71	-0.01
Aroclor-1242-3 (3)	5.68	5.68	5.58	5.78	0.00
Aroclor-1242-4 (4)	5.78	5.78	5.68	5.88	0.00
Aroclor-1242-5 (5)	6.54	6.53	6.43	6.63	0.00
Tetrachloro-m-xylene	4.41	4.41	4.31	4.51	0.00
Decachlorobiphenyl	10.35	10.34	10.24	10.44	-0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/08/2023 Initial Calibration Date(s): 08/07/2023 08/07/2023

Continuing Calib Time: 21:07 Initial Calibration Time(s): 16:22 23:59

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1242-1 (1)	4.71	4.71	4.61	4.81	0.00
Aroclor-1242-2 (2)	4.73	4.73	4.63	4.83	0.01
Aroclor-1242-3 (3)	4.90	4.90	4.80	5.00	0.00
Aroclor-1242-4 (4)	4.99	4.99	4.89	5.09	0.00
Aroclor-1242-5 (5)	5.51	5.51	5.41	5.61	0.00
Tetrachloro-m-xylene	3.62	3.62	3.52	3.72	0.00
Decachlorobiphenyl	8.64	8.65	8.55	8.75	0.01