



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/08/2023 Initial Calibration Date(s): 08/07/2023 08/07/2023

Continuing Calib Time: 21:41 Initial Calibration Time(s): 16:22 23:59

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|-------|------------|
| Aroclor-1254-1       | (1) | 6.47       | 6.46      | 6.36                   | 6.56  | -0.01      |
| Aroclor-1254-2       | (2) | 6.69       | 6.69      | 6.59                   | 6.79  | 0.00       |
| Aroclor-1254-3       | (3) | 7.07       | 7.06      | 6.96                   | 7.16  | -0.01      |
| Aroclor-1254-4       | (4) | 7.36       | 7.35      | 7.25                   | 7.45  | -0.01      |
| Aroclor-1254-5       | (5) | 7.79       | 7.79      | 7.69                   | 7.89  | 0.00       |
| Tetrachloro-m-xylene |     | 4.41       | 4.41      | 4.31                   | 4.51  | 0.00       |
| Decachlorobiphenyl   |     | 10.35      | 10.34     | 10.24                  | 10.44 | -0.01      |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/08/2023 Initial Calibration Date(s): 08/07/2023 08/07/2023

Continuing Calib Time: 21:41 Initial Calibration Time(s): 16:22 23:59

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.51       | 5.51      | 5.41      | 5.61 | 0.00       |
| Aroclor-1254-2 (2)   | 5.66       | 5.66      | 5.56      | 5.76 | 0.00       |
| Aroclor-1254-3 (3)   | 6.06       | 6.06      | 5.96      | 6.16 | 0.00       |
| Aroclor-1254-4 (4)   | 6.29       | 6.29      | 6.19      | 6.39 | 0.00       |
| Aroclor-1254-5 (5)   | 6.71       | 6.71      | 6.61      | 6.81 | 0.00       |
| Tetrachloro-m-xylene | 3.62       | 3.62      | 3.52      | 3.72 | 0.00       |
| Decachlorobiphenyl   | 8.64       | 8.64      | 8.54      | 8.74 | 0.00       |