



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/09/2023 Initial Calibration Date(s): 08/07/2023 08/07/2023

Continuing Calib Time: 10:21 Initial Calibration Time(s): 16:22 23:59

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1016-1 (1)	5.60	5.59	5.49	5.69	0.00
Aroclor-1016-2 (2)	5.62	5.61	5.51	5.71	-0.01
Aroclor-1016-3 (3)	5.68	5.68	5.58	5.78	0.00
Aroclor-1016-4 (4)	5.78	5.78	5.68	5.88	0.00
Aroclor-1016-5 (5)	6.08	6.08	5.98	6.18	0.00
Aroclor-1260-1 (1)	7.24	7.23	7.13	7.33	-0.01
Aroclor-1260-2 (2)	7.50	7.49	7.39	7.59	-0.01
Aroclor-1260-3 (3)	7.87	7.86	7.76	7.96	-0.01
Aroclor-1260-4 (4)	8.10	8.10	8.00	8.20	0.00
Aroclor-1260-5 (5)	8.44	8.43	8.33	8.53	-0.01
Tetrachloro-m-xylene	4.41	4.41	4.31	4.51	0.00
Decachlorobiphenyl	10.35	10.34	10.24	10.44	-0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/09/2023 Initial Calibration Date(s): 08/07/2023 08/07/2023

Continuing Calib Time: 10:21 Initial Calibration Time(s): 16:22 23:59

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1016-1 (1)	4.71	4.71	4.61	4.81	0.01
Aroclor-1016-2 (2)	4.72	4.73	4.63	4.83	0.01
Aroclor-1016-3 (3)	4.90	4.90	4.80	5.00	0.00
Aroclor-1016-4 (4)	4.94	4.95	4.85	5.05	0.01
Aroclor-1016-5 (5)	5.16	5.16	5.06	5.26	0.00
Aroclor-1260-1 (1)	6.19	6.20	6.10	6.30	0.01
Aroclor-1260-2 (2)	6.38	6.39	6.29	6.49	0.01
Aroclor-1260-3 (3)	6.54	6.54	6.44	6.64	0.00
Aroclor-1260-4 (4)	7.01	7.01	6.91	7.11	0.00
Aroclor-1260-5 (5)	7.25	7.26	7.16	7.36	0.01
Tetrachloro-m-xylene	3.62	3.62	3.52	3.72	0.00
Decachlorobiphenyl	8.64	8.64	8.54	8.74	0.00