



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/10/2023 Initial Calibration Date(s): 08/07/2023 08/07/2023

Continuing Calib Time: 05:13 Initial Calibration Time(s): 16:22 23:59

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	6.47	6.46	6.36	6.56	-0.01
Aroclor-1254-2 (2)	6.69	6.69	6.59	6.79	0.00
Aroclor-1254-3 (3)	7.07	7.06	6.96	7.16	-0.01
Aroclor-1254-4 (4)	7.36	7.35	7.25	7.45	-0.01
Aroclor-1254-5 (5)	7.79	7.79	7.69	7.89	0.00
Tetrachloro-m-xylene	4.41	4.41	4.31	4.51	0.00
Decachlorobiphenyl	10.35	10.34	10.24	10.44	-0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/10/2023 Initial Calibration Date(s): 08/07/2023 08/07/2023

Continuing Calib Time: 05:13 Initial Calibration Time(s): 16:22 23:59

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.51	5.51	5.41	5.61	0.00
Aroclor-1254-2 (2)	5.66	5.66	5.56	5.76	0.00
Aroclor-1254-3 (3)	6.06	6.06	5.96	6.16	0.00
Aroclor-1254-4 (4)	6.29	6.29	6.19	6.39	0.00
Aroclor-1254-5 (5)	6.71	6.71	6.61	6.81	0.00
Tetrachloro-m-xylene	3.62	3.62	3.52	3.72	0.00
Decachlorobiphenyl	8.64	8.64	8.54	8.74	0.00