



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/09/2023 Initial Calibration Date(s): 08/08/2023 08/08/2023

Continuing Calib Time: 10:34 Initial Calibration Time(s): 18:17 21:16

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.59	5.59	5.49	5.69	0.00
Aroclor-1016-2	(2)	5.61	5.62	5.52	5.72	0.01
Aroclor-1016-3	(3)	5.68	5.68	5.58	5.78	0.00
Aroclor-1016-4	(4)	5.78	5.78	5.68	5.88	0.00
Aroclor-1016-5	(5)	6.07	6.08	5.98	6.18	0.01
Aroclor-1260-1	(1)	7.20	7.21	7.11	7.31	0.01
Aroclor-1260-2	(2)	7.46	7.46	7.36	7.56	0.00
Aroclor-1260-3	(3)	7.82	7.83	7.73	7.93	0.01
Aroclor-1260-4	(4)	8.05	8.05	7.95	8.15	0.00
Aroclor-1260-5	(5)	8.37	8.37	8.27	8.47	0.00
Tetrachloro-m-xylene		4.42	4.42	4.32	4.52	0.01
Decachlorobiphenyl		10.21	10.21	10.11	10.31	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/09/2023 Initial Calibration Date(s): 08/08/2023 08/08/2023

Continuing Calib Time: 10:34 Initial Calibration Time(s): 18:17 21:16

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1016-1 (1)	4.74	4.74	4.64	4.84	0.00
Aroclor-1016-2 (2)	4.76	4.76	4.66	4.86	0.00
Aroclor-1016-3 (3)	4.94	4.94	4.84	5.04	0.01
Aroclor-1016-4 (4)	4.98	4.98	4.88	5.08	0.00
Aroclor-1016-5 (5)	5.19	5.20	5.10	5.30	0.01
Aroclor-1260-1 (1)	6.23	6.23	6.13	6.33	0.00
Aroclor-1260-2 (2)	6.42	6.42	6.32	6.52	0.00
Aroclor-1260-3 (3)	6.57	6.58	6.48	6.68	0.01
Aroclor-1260-4 (4)	7.05	7.05	6.95	7.15	0.00
Aroclor-1260-5 (5)	7.29	7.29	7.19	7.39	0.00
Tetrachloro-m-xylene	3.65	3.65	3.55	3.75	0.00
Decachlorobiphenyl	8.68	8.69	8.59	8.79	0.01