



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/09/2023 Initial Calibration Date(s): 08/08/2023 08/08/2023

Continuing Calib Time: 12:15 Initial Calibration Time(s): 16:30 19:12

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.92	4.91	4.81	5.01	-0.01
Aroclor-1016-2	(2)	4.94	4.94	4.84	5.04	0.00
Aroclor-1016-3	(3)	5.01	5.00	4.90	5.10	-0.01
Aroclor-1016-4	(4)	5.11	5.10	5.00	5.20	-0.01
Aroclor-1016-5	(5)	5.43	5.42	5.32	5.52	-0.01
Aroclor-1260-1	(1)	6.65	6.64	6.54	6.74	0.00
Aroclor-1260-2	(2)	6.93	6.92	6.82	7.02	-0.01
Aroclor-1260-3	(3)	7.32	7.31	7.21	7.41	-0.01
Aroclor-1260-4	(4)	7.56	7.55	7.45	7.65	-0.01
Aroclor-1260-5	(5)	7.90	7.89	7.79	7.99	-0.01
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		9.63	9.62	9.52	9.72	-0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/09/2023 Initial Calibration Date(s): 08/08/2023 08/08/2023

Continuing Calib Time: 12:15 Initial Calibration Time(s): 16:30 19:12

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.08	4.08	3.98	4.18	0.00
Aroclor-1016-2	(2)	4.10	4.10	4.00	4.20	0.00
Aroclor-1016-3	(3)	4.28	4.28	4.18	4.38	0.00
Aroclor-1016-4	(4)	4.33	4.33	4.23	4.43	0.00
Aroclor-1016-5	(5)	4.55	4.55	4.45	4.65	0.00
Aroclor-1260-1	(1)	5.64	5.64	5.54	5.74	0.00
Aroclor-1260-2	(2)	5.85	5.85	5.75	5.95	0.00
Aroclor-1260-3	(3)	6.01	6.01	5.91	6.11	0.00
Aroclor-1260-4	(4)	6.52	6.52	6.42	6.62	0.00
Aroclor-1260-5	(5)	6.78	6.78	6.68	6.88	0.00
Tetrachloro-m-xylene		2.96	2.96	2.86	3.06	0.01
Decachlorobiphenyl		8.24	8.24	8.14	8.34	0.00