



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/09/2023 Initial Calibration Date(s): 08/08/2023 08/08/2023

Continuing Calib Time: 14:57 Initial Calibration Time(s): 16:30 19:12

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1016-1 (1)	4.91	4.91	4.81	5.01	0.00
Aroclor-1016-2 (2)	4.94	4.94	4.84	5.04	0.00
Aroclor-1016-3 (3)	5.00	5.00	4.90	5.10	0.00
Aroclor-1016-4 (4)	5.10	5.10	5.00	5.20	0.00
Aroclor-1016-5 (5)	5.42	5.42	5.32	5.52	0.00
Aroclor-1260-1 (1)	6.64	6.64	6.54	6.74	0.00
Aroclor-1260-2 (2)	6.92	6.92	6.82	7.02	0.00
Aroclor-1260-3 (3)	7.31	7.31	7.21	7.41	0.00
Aroclor-1260-4 (4)	7.55	7.55	7.45	7.65	0.00
Aroclor-1260-5 (5)	7.89	7.89	7.79	7.99	0.00
Tetrachloro-m-xylene	3.68	3.69	3.59	3.79	0.01
Decachlorobiphenyl	9.62	9.62	9.52	9.72	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Continuing Calib Date: 08/09/2023 Initial Calibration Date(s): 08/08/2023 08/08/2023

Continuing Calib Time: 14:57 Initial Calibration Time(s): 16:30 19:12

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1016-1 (1)	4.08	4.08	3.98	4.18	0.00
Aroclor-1016-2 (2)	4.10	4.10	4.00	4.20	0.00
Aroclor-1016-3 (3)	4.28	4.28	4.18	4.38	0.00
Aroclor-1016-4 (4)	4.33	4.33	4.23	4.43	0.00
Aroclor-1016-5 (5)	4.55	4.55	4.45	4.65	0.00
Aroclor-1260-1 (1)	5.64	5.64	5.54	5.74	0.00
Aroclor-1260-2 (2)	5.85	5.85	5.75	5.95	0.00
Aroclor-1260-3 (3)	6.01	6.01	5.91	6.11	0.00
Aroclor-1260-4 (4)	6.52	6.52	6.42	6.62	0.01
Aroclor-1260-5 (5)	6.78	6.78	6.68	6.88	0.00
Tetrachloro-m-xylene	2.96	2.96	2.86	3.06	0.01
Decachlorobiphenyl	8.24	8.24	8.14	8.34	0.00