



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 08/07/2023 08/07/2023

Client Sample No.: CCAL02 Date Analyzed: 08/09/2023

Lab Sample No.: AR1660CCC500 Data File : PO096830.D Time Analyzed: 10:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.595	5.491	5.691	528.530	500.000	5.7
Aroclor-1016-2	5.619	5.514	5.714	527.240	500.000	5.4
Aroclor-1016-3	5.681	5.577	5.777	510.310	500.000	2.1
Aroclor-1016-4	5.781	5.676	5.876	527.890	500.000	5.6
Aroclor-1016-5	6.083	5.977	6.177	540.520	500.000	8.1
Aroclor-1260-1	7.238	7.130	7.330	513.490	500.000	2.7
Aroclor-1260-2	7.502	7.393	7.593	528.790	500.000	5.8
Aroclor-1260-3	7.871	7.764	7.964	538.040	500.000	7.6
Aroclor-1260-4	8.104	7.995	8.195	542.650	500.000	8.5
Aroclor-1260-5	8.439	8.329	8.529	530.660	500.000	6.1
Decachlorobiphenyl	10.351	10.240	10.440	54.030	50.000	8.1
Tetrachloro-m-xylene	4.410	4.307	4.507	53.500	50.000	7.0



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 08/07/2023 08/07/2023

Client Sample No.: CCAL02 Date Analyzed: 08/09/2023

Lab Sample No.: AR1660CCC500 Data File : PO096830.D Time Analyzed: 10:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.705	4.608	4.808	523.830	500.000	4.8
Aroclor-1016-2	4.724	4.626	4.826	520.020	500.000	4.0
Aroclor-1016-3	4.900	4.803	5.003	544.680	500.000	8.9
Aroclor-1016-4	4.943	4.846	5.046	517.030	500.000	3.4
Aroclor-1016-5	5.156	5.060	5.260	506.110	500.000	1.2
Aroclor-1260-1	6.194	6.096	6.296	526.240	500.000	5.2
Aroclor-1260-2	6.383	6.285	6.485	516.010	500.000	3.2
Aroclor-1260-3	6.536	6.439	6.639	516.290	500.000	3.3
Aroclor-1260-4	7.010	6.912	7.112	518.600	500.000	3.7
Aroclor-1260-5	7.253	7.155	7.355	517.580	500.000	3.5
Decachlorobiphenyl	8.641	8.544	8.744	52.590	50.000	5.2
Tetrachloro-m-xylene	3.619	3.521	3.721	53.890	50.000	7.8