



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 08/08/2023 08/08/2023

Client Sample No.: CCAL01 Date Analyzed: 08/09/2023

Lab Sample No.: AR1660CCC500 Data File : PP059257.D Time Analyzed: 10:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.592	5.494	5.694	492.920	500.000	-1.4
Aroclor-1016-2	5.614	5.516	5.716	495.440	500.000	-0.9
Aroclor-1016-3	5.677	5.579	5.779	501.340	500.000	0.3
Aroclor-1016-4	5.776	5.678	5.878	508.170	500.000	1.6
Aroclor-1016-5	6.073	5.975	6.175	512.390	500.000	2.5
Aroclor-1260-1	7.204	7.107	7.307	532.720	500.000	6.5
Aroclor-1260-2	7.462	7.364	7.564	510.720	500.000	2.1
Aroclor-1260-3	7.823	7.726	7.926	564.680	500.000	12.9
Aroclor-1260-4	8.049	7.951	8.151	531.180	500.000	6.2
Aroclor-1260-5	8.370	8.274	8.474	512.870	500.000	2.6
Decachlorobiphenyl	10.209	10.111	10.311	51.640	50.000	3.3
Tetrachloro-m-xylene	4.415	4.316	4.516	49.160	50.000	-1.7



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 08/08/2023 08/08/2023

Client Sample No.: CCAL01 Date Analyzed: 08/09/2023

Lab Sample No.: AR1660CCC500 Data File : PP059257.D Time Analyzed: 10:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.740	4.643	4.843	508.510	500.000	1.7
Aroclor-1016-2	4.759	4.662	4.862	518.800	500.000	3.8
Aroclor-1016-3	4.935	4.838	5.038	531.500	500.000	6.3
Aroclor-1016-4	4.978	4.881	5.081	515.550	500.000	3.1
Aroclor-1016-5	5.191	5.095	5.295	514.490	500.000	2.9
Aroclor-1260-1	6.230	6.133	6.333	518.100	500.000	3.6
Aroclor-1260-2	6.419	6.323	6.523	523.540	500.000	4.7
Aroclor-1260-3	6.573	6.477	6.677	508.640	500.000	1.7
Aroclor-1260-4	7.047	6.950	7.150	529.750	500.000	6.0
Aroclor-1260-5	7.290	7.194	7.394	533.760	500.000	6.8
Decachlorobiphenyl	8.684	8.589	8.789	52.190	50.000	4.4
Tetrachloro-m-xylene	3.647	3.549	3.749	50.900	50.000	1.8