



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 08/08/2023 08/08/2023

Client Sample No.: CCAL02 Date Analyzed: 08/09/2023

Lab Sample No.: AR1660CCC500 Data File : PP059269.D Time Analyzed: 16:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.593	5.494	5.694	476.900	500.000	-4.6
Aroclor-1016-2	5.615	5.516	5.716	472.860	500.000	-5.4
Aroclor-1016-3	5.677	5.579	5.779	480.850	500.000	-3.8
Aroclor-1016-4	5.776	5.678	5.878	488.070	500.000	-2.4
Aroclor-1016-5	6.073	5.975	6.175	491.140	500.000	-1.8
Aroclor-1260-1	7.205	7.107	7.307	503.250	500.000	0.7
Aroclor-1260-2	7.463	7.364	7.564	488.800	500.000	-2.2
Aroclor-1260-3	7.824	7.726	7.926	540.970	500.000	8.2
Aroclor-1260-4	8.050	7.951	8.151	491.020	500.000	-1.8
Aroclor-1260-5	8.372	8.274	8.474	487.530	500.000	-2.5
Decachlorobiphenyl	10.210	10.111	10.311	49.080	50.000	-1.8
Tetrachloro-m-xylene	4.416	4.316	4.516	46.860	50.000	-6.3



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 08/08/2023 08/08/2023

Client Sample No.: CCAL02 Date Analyzed: 08/09/2023

Lab Sample No.: AR1660CCC500 Data File : PP059269.D Time Analyzed: 16:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.740	4.643	4.843	488.120	500.000	-2.4
Aroclor-1016-2	4.759	4.662	4.862	496.830	500.000	-0.6
Aroclor-1016-3	4.935	4.838	5.038	509.350	500.000	1.9
Aroclor-1016-4	4.978	4.881	5.081	497.880	500.000	-0.4
Aroclor-1016-5	5.192	5.095	5.295	497.530	500.000	-0.5
Aroclor-1260-1	6.230	6.133	6.333	495.360	500.000	-0.9
Aroclor-1260-2	6.420	6.323	6.523	500.680	500.000	0.1
Aroclor-1260-3	6.574	6.477	6.677	486.000	500.000	-2.8
Aroclor-1260-4	7.047	6.950	7.150	506.620	500.000	1.3
Aroclor-1260-5	7.291	7.194	7.394	512.200	500.000	2.4
Decachlorobiphenyl	8.684	8.589	8.789	50.600	50.000	1.2
Tetrachloro-m-xylene	3.647	3.549	3.749	48.860	50.000	-2.3