



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 08/08/2023 08/08/2023

Client Sample No.: CCAL01 Date Analyzed: 08/09/2023

Lab Sample No.: AR1660CCC500 Data File : PR062551.D Time Analyzed: 09:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.915	4.814	5.014	498.240	500.000	-0.4
Aroclor-1016-2	4.936	4.837	5.037	502.410	500.000	0.5
Aroclor-1016-3	5.001	4.901	5.101	496.830	500.000	-0.6
Aroclor-1016-4	5.104	5.004	5.204	508.920	500.000	1.8
Aroclor-1016-5	5.420	5.321	5.521	503.010	500.000	0.6
Aroclor-1260-1	6.637	6.538	6.738	487.430	500.000	-2.5
Aroclor-1260-2	6.921	6.819	7.019	502.370	500.000	0.5
Aroclor-1260-3	7.311	7.211	7.411	514.710	500.000	2.9
Aroclor-1260-4	7.553	7.453	7.653	518.270	500.000	3.7
Aroclor-1260-5	7.892	7.792	7.992	510.530	500.000	2.1
Decachlorobiphenyl	9.619	9.517	9.717	50.650	50.000	1.3
Tetrachloro-m-xylene	3.684	3.585	3.785	49.650	50.000	-0.7



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 08/08/2023 08/08/2023

Client Sample No.: CCAL01 Date Analyzed: 08/09/2023

Lab Sample No.: AR1660CCC500 Data File : PR062551.D Time Analyzed: 09:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.078	3.979	4.179	486.260	500.000	-2.7
Aroclor-1016-2	4.096	3.998	4.198	491.050	500.000	-1.8
Aroclor-1016-3	4.276	4.177	4.377	495.200	500.000	-1.0
Aroclor-1016-4	4.327	4.228	4.428	490.350	500.000	-1.9
Aroclor-1016-5	4.545	4.447	4.647	473.260	500.000	-5.3
Aroclor-1260-1	5.641	5.543	5.743	491.320	500.000	-1.7
Aroclor-1260-2	5.847	5.748	5.948	490.370	500.000	-1.9
Aroclor-1260-3	6.007	5.908	6.108	494.150	500.000	-1.2
Aroclor-1260-4	6.514	6.416	6.616	490.140	500.000	-2.0
Aroclor-1260-5	6.780	6.681	6.881	496.380	500.000	-0.7
Decachlorobiphenyl	8.236	8.137	8.337	47.420	50.000	-5.2
Tetrachloro-m-xylene	2.955	2.856	3.056	50.190	50.000	0.4