



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 08/08/2023 08/08/2023

Client Sample No.: CCAL02 Date Analyzed: 08/09/2023

Lab Sample No.: AR1660CCC500 Data File : PR062562.D Time Analyzed: 12:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.922	4.814	5.014	493.500	500.000	-1.3
Aroclor-1016-2	4.944	4.837	5.037	494.060	500.000	-1.2
Aroclor-1016-3	5.009	4.901	5.101	491.260	500.000	-1.7
Aroclor-1016-4	5.112	5.004	5.204	497.730	500.000	-0.5
Aroclor-1016-5	5.429	5.321	5.521	493.430	500.000	-1.3
Aroclor-1260-1	6.645	6.538	6.738	491.160	500.000	-1.8
Aroclor-1260-2	6.929	6.819	7.019	494.730	500.000	-1.1
Aroclor-1260-3	7.319	7.211	7.411	499.460	500.000	-0.1
Aroclor-1260-4	7.562	7.453	7.653	506.420	500.000	1.3
Aroclor-1260-5	7.901	7.792	7.992	500.220	500.000	0.0
Decachlorobiphenyl	9.629	9.517	9.717	47.980	50.000	-4.0
Tetrachloro-m-xylene	3.692	3.585	3.785	49.110	50.000	-1.8



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 08/08/2023 08/08/2023

Client Sample No.: CCAL02 Date Analyzed: 08/09/2023

Lab Sample No.: AR1660CCC500 Data File : PR062562.D Time Analyzed: 12:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.080	3.979	4.179	485.240	500.000	-3.0
Aroclor-1016-2	4.098	3.998	4.198	486.820	500.000	-2.6
Aroclor-1016-3	4.278	4.177	4.377	490.280	500.000	-1.9
Aroclor-1016-4	4.328	4.228	4.428	487.940	500.000	-2.4
Aroclor-1016-5	4.547	4.447	4.647	475.510	500.000	-4.9
Aroclor-1260-1	5.644	5.543	5.743	493.790	500.000	-1.2
Aroclor-1260-2	5.849	5.748	5.948	495.950	500.000	-0.8
Aroclor-1260-3	6.009	5.908	6.108	499.270	500.000	-0.1
Aroclor-1260-4	6.518	6.416	6.616	494.160	500.000	-1.2
Aroclor-1260-5	6.783	6.681	6.881	498.460	500.000	-0.3
Decachlorobiphenyl	8.242	8.137	8.337	48.450	50.000	-3.1
Tetrachloro-m-xylene	2.955	2.856	3.056	50.260	50.000	0.5