



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Instrument ID: ECD_D Date(s) Analyzed: 08/03/2023 08/03/2023

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.25	6.15	6.35	24321700
		2	7.16	7.06	7.26	59365600
		3	7.26	7.16	7.36	35958400
		4	7.59	7.49	7.69	44341100
		5	7.95	7.85	8.05	44274000



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Instrument ID: ECD_D Date(s) Analyzed: 08/03/2023 08/03/2023

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	4.99	4.89	5.09	9129920
		2	5.67	5.57	5.77	9177080
		3	6.59	6.49	6.69	27065200
		4	6.72	6.62	6.82	42130200
		5	7.03	6.93	7.13	15429700