



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Instrument ID: ECD_L Date(s) Analyzed: 08/03/2023 08/03/2023

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.19	6.09	6.29	19868000
		2	7.01	6.91	7.11	64597700
		3	7.10	7.00	7.20	45510700
		4	7.20	7.10	7.30	26554600
		5	7.53	7.43	7.63	33541500



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG NO.: O3572

Instrument ID: ECD_L Date(s) Analyzed: 08/03/2023 08/03/2023

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.25	5.15	5.35	6819470
		2	5.42	5.32	5.52	5574820
		3	5.90	5.80	6.00	8998830
		4	6.53	6.43	6.63	30747200
		5	6.66	6.56	6.76	41768600