



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
ProjectID: LOD-LOQ Study
Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG No.: O3572

Calibration Sequence : FE083023		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	231692497	136290	FE044034.D
850	127176069	149619	FE044035.D
340	51813025	152391	FE044036.D
170	25668447	150991	FE044037.D
85	13632514	160383	FE044038.D
AVG RF : 149935		% RSD : 5.801	AVG RT : 15.034