



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
ProjectID: LOD-LOQ Study
Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG No.: O3572

Calibration Sequence : FF080223		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	191682712	112755	FF013110.D
850	96982089	114097	FF013111.D
340	38408067	112965	FF013112.D
170	17706145	104154	FF013113.D
85	9614410	113111	FF013114.D
AVG RF : 111416		% RSD : 3.673	AVG RT : 14.9874