



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
ProjectID: LOD-LOQ Study
Lab Code: CHEM Case No.: O3572 SAS No.: O3572 SDG No.: O3572

Calibration Sequence : FG080223		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	189624312	111544	FG012605.D
850	96544512	113582	FG012606.D
340	38331130	112739	FG012607.D
170	19652426	115603	FG012608.D
85	10116636	119019	FG012609.D
AVG RF : 114497		% RSD : 2.558	AVG RT : 15.0778