

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>CHEMTECH</u>	Contract: <u>LABE01</u>
Lab Code: <u>CHEM</u> Case No.: <u>O3631</u>	SAS No.: <u>O3631</u> SDG No.: <u>O3631</u>
Instrument ID: <u>MSVOA_N</u>	Calibration Date(s): <u>07/10/2023</u> <u>07/10/2023</u>
Heated Purge: (Y/N) <u>N</u>	Calibration Time(s): <u>12:21</u> <u>14:46</u>
GC Column: <u>RXI-624</u> ID: <u>0.25</u> (mm)	

LAB FILE ID:	RRF100 = VN078449.D	RRF050 = VN078450.D	RRF040 = VN078451.D					
	RRF020 = VN078452.D	RRF005 = VN078453.D	RRF001 = VN078454.D					
COMPOUND	RRF100	RRF050	RRF040	RRF020	RRF005	RRF001	RRF	% RSD
Dichlorodifluoromethane	0.565	0.586	0.609	0.576	0.691	0.650	0.613	7.9
Chloromethane	0.566	0.614	0.651	0.617	0.733	0.818	0.666	13.9
Vinyl Chloride	0.802	0.813	0.805	0.842	0.936	0.974	0.862	8.7
Bromomethane	0.509	0.569	0.586	0.620	0.795		0.616	17.5
Chloroethane	0.500	0.570	0.523	0.560	0.592	0.751	0.583	15.2
Trichlorofluoromethane	0.945	0.983	0.993	1.010	1.108	1.143	1.030	7.5
1,1,2-Trichlorotrifluoroethane	0.513	0.537	0.542	0.533	0.577	0.569	0.545	4.4
1,1-Dichloroethene	0.502	0.519	0.523	0.509	0.567	0.545	0.528	4.6
Acetone	0.187	0.198	0.207	0.200	0.224	0.237	0.209	8.8
Carbon Disulfide	1.407	1.440	1.438	1.443	1.615	1.810	1.526	10.3
Methyl tert-butyl Ether	1.689	1.768	1.788	1.704	1.781	1.701	1.738	2.6
Methyl Acetate	0.796	0.865	0.866	0.829	0.933	1.015	0.884	8.9
Methylene Chloride	0.585	0.615	0.625	0.603	0.676	0.941	0.674	20
trans-1,2-Dichloroethene	0.562	0.592	0.591	0.581	0.631	0.610	0.595	4
1,1-Dichloroethane	1.032	1.076	1.062	1.056	1.156	1.144	1.088	4.6
Cyclohexane	0.825	0.830	0.882	0.876	1.117		0.906	13.3
2-Butanone	0.314	0.330	0.341	0.326	0.329	0.332	0.329	2.7
Carbon Tetrachloride	0.537	0.526	0.545	0.531	0.518	0.539	0.533	1.8
cis-1,2-Dichloroethene	0.673	0.688	0.696	0.681	0.711	0.678	0.688	2
Bromochloromethane	0.493	0.513	0.504	0.504	0.492	0.567	0.512	5.5
Chloroform	1.098	1.146	1.155	1.118	1.157	1.122	1.132	2.1
1,1,1-Trichloroethane	0.993	1.033	1.043	1.017	1.058	0.948	1.015	3.9
Methylcyclohexane	0.464	0.446	0.453	0.442	0.423	0.455	0.447	3.1
Benzene	1.458	1.443	1.478	1.463	1.450	1.420	1.452	1.4
1,2-Dichloroethane	0.485	0.475	0.487	0.497	0.497	0.522	0.494	3.3
Trichloroethene	0.382	0.371	0.387	0.376	0.373	0.458	0.391	8.5
1,2-Dichloropropane	0.372	0.358	0.368	0.374	0.367	0.367	0.368	1.5
Bromodichloromethane	0.526	0.518	0.522	0.526	0.503	0.518	0.519	1.7
4-Methyl-2-Pentanone	0.410	0.410	0.427	0.420	0.388	0.346	0.400	7.4
Toluene	0.930	0.902	0.936	0.878	0.869	0.885	0.900	3.1
t-1,3-Dichloropropene	0.570	0.543	0.557	0.531	0.513	0.476	0.532	6.3
cis-1,3-Dichloropropene	0.611	0.591	0.601	0.578	0.551	0.467	0.567	9.3
1,1,2-Trichloroethane	0.356	0.351	0.360	0.359	0.348	0.353	0.355	1.3

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>CHEMTECH</u>	Contract: <u>LABE01</u>
Lab Code: <u>CHEM</u> Case No.: <u>O3631</u>	SAS No.: <u>O3631</u> SDG No.: <u>O3631</u>
Instrument ID: <u>MSVOA_N</u>	Calibration Date(s): <u>07/10/2023</u> <u>07/10/2023</u>
Heated Purge: (Y/N) <u>N</u>	Calibration Time(s): <u>12:21</u> <u>14:46</u>
GC Column: <u>RXI-624</u> ID: <u>0.25</u> (mm)	

LAB FILE ID:		RRF100 = VN078449.D		RRF050 = VN078450.D		RRF040 = VN078451.D		
		RRF020 = VN078452.D		RRF005 = VN078453.D		RRF001 = VN078454.D		
COMPOUND	RRF100	RRF050	RRF040	RRF020	RRF005	RRF001	RRF	% RSD
2-Hexanone	0.293	0.298	0.306	0.291	0.282	0.237	0.284	8.7
Dibromochloromethane	0.412	0.401	0.406	0.393	0.377	0.360	0.391	5
1,2-Dibromoethane	0.373	0.368	0.381	0.366	0.368	0.335	0.365	4.3
Tetrachloroethene	0.366	0.373	0.389	0.396	0.391	0.340	0.376	5.5
Chlorobenzene	1.094	1.095	1.124	1.112	1.139	1.142	1.118	1.9
Ethyl Benzene	1.974	1.945	1.965	1.929	1.860	1.662	1.889	6.3
m/p-Xylenes	0.747	0.737	0.745	0.721	0.704	0.536	0.698	11.7
o-Xylene	0.736	0.718	0.725	0.713	0.686	0.654	0.705	4.3
Styrene	1.222	1.170	1.190	1.117	1.056	0.867	1.104	11.8
Bromoform	0.304	0.306	0.311	0.294	0.297	0.236	0.291	9.5
Isopropylbenzene	4.223	4.339	4.344	4.665	5.480	5.138	4.698	10.8
1,1,2,2-Tetrachloroethane	1.245	1.370	1.396	1.498	1.856	2.152	1.586	21.8
n-propylbenzene	4.818	4.824	4.885	5.017	5.527	5.076	5.024	5.3
1,3,5-Trimethylbenzene	3.386	3.486	3.562	3.773	4.124	3.941	3.712	7.7
tert-Butylbenzene	2.856	2.928	2.931	3.159	3.620	3.415	3.152	9.8
1,2,4-Trimethylbenzene	3.353	3.440	3.566	3.649	4.067	3.560	3.606	6.9
sec-Butylbenzene	3.700	3.733	3.741	3.923	4.614	4.191	3.984	9
p-Isopropyltoluene	3.051	3.079	3.135	3.204	3.565	3.159	3.199	5.9
1,3-Dichlorobenzene	1.729	1.746	1.758	1.767	1.968	1.855	1.804	5.1
1,4-Dichlorobenzene	1.688	1.723	1.710	1.775	1.939	2.064	1.817	8.3
n-Butylbenzene	2.310	2.322	2.343	2.296	2.555	1.906	2.289	9.2
1,2-Dichlorobenzene	1.673	1.724	1.730	1.808	1.985	1.849	1.795	6.3
1,2-Dibromo-3-Chloropropane	0.218	0.227	0.236	0.246	0.277	0.365	0.262	20.8
1,2,4-Trichlorobenzene	0.630	0.609	0.586	0.555	0.512	0.545	0.573	7.6
1,2,3-Trichlorobenzene	0.668	0.660	0.658	0.646	0.627	0.694	0.659	3.4
1,2-Dichloroethane-d4	0.652	0.662	0.657	0.675	0.706		0.670	3.2
Dibromofluoromethane	0.338	0.331	0.330	0.347	0.351		0.339	2.8
Toluene-d8	1.319	1.251	1.247	1.281	1.313		1.282	2.6
4-Bromofluorobenzene	0.430	0.397	0.392	0.388	0.368		0.395	5.7

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.