

Surrogate Summary

SDG No.: O3631

Client: LaBella Associates P.C.

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
O3631-01	SB-02(2-4)	1,2-Dichloroethane-d4	50	46.1	92	50	163
		Dibromofluoromethane	50	47.6	95	54	147
		Toluene-d8	50	49.9	100	58	134
		4-Bromofluorobenzene	50	45.4	91	30	143
O3631-02MS	SB-02(2-4)MS	1,2-Dichloroethane-d4	50	29.9	60	50	163
		Dibromofluoromethane	50	41.3	83	54	147
		Toluene-d8	50	43.1	86	58	134
		4-Bromofluorobenzene	50	33.7	67	30	143
O3631-03MSD	SB-02(2-4)MSD	1,2-Dichloroethane-d4	50	52.6	105	50	163
		Dibromofluoromethane	50	53.6	107	54	147
		Toluene-d8	50	50.3	101	58	134
		4-Bromofluorobenzene	50	43.9	88	30	143
O3631-04	SB-04(2-4)	1,2-Dichloroethane-d4	50	58.1	116	50	163
		Dibromofluoromethane	50	50.2	100	54	147
		Toluene-d8	50	50.8	102	58	134
		4-Bromofluorobenzene	50	50.9	102	30	143
O3631-05	SB-07(3-5)	1,2-Dichloroethane-d4	50	56.3	113	50	163
		Dibromofluoromethane	50	50.7	101	54	147
		Toluene-d8	50	50.4	101	58	134
		4-Bromofluorobenzene	50	50.6	101	30	143
O3631-06	SB-09(3-5)	1,2-Dichloroethane-d4	50	52.9	106	50	163
		Dibromofluoromethane	50	48.9	98	54	147
		Toluene-d8	50	50.7	101	58	134
		4-Bromofluorobenzene	50	49.0	98	30	143
O3631-07	SB-10(5-6)	1,2-Dichloroethane-d4	50	55.0	110	50	163
		Dibromofluoromethane	50	49.5	99	54	147
		Toluene-d8	50	50.3	100	58	134
		4-Bromofluorobenzene	50	48.2	96	30	143
O3631-08	SB-11(3.5-4.5)	1,2-Dichloroethane-d4	50	53.7	107	50	163
		Dibromofluoromethane	50	49.7	99	54	147
		Toluene-d8	50	50.7	101	58	134
		4-Bromofluorobenzene	50	50.0	100	30	143
O3631-09	DUP	1,2-Dichloroethane-d4	50	57.7	115	50	163
		Dibromofluoromethane	50	50.8	102	54	147
		Toluene-d8	50	51.0	102	58	134
		4-Bromofluorobenzene	50	51.3	102	30	143
VY0717SBL01	VY0717SBL01	1,2-Dichloroethane-d4	50	52.5	105	50	163
		Dibromofluoromethane	50	49.6	99	54	147
		Toluene-d8	50	50.0	100	58	134
		4-Bromofluorobenzene	50	49.0	98	30	143
VY0717SBS01	VY0717SBS01	1,2-Dichloroethane-d4	50	51.3	103	50	163
		Dibromofluoromethane	50	50.0	100	54	147
		Toluene-d8	50	49.6	99	58	134
		4-Bromofluorobenzene	50	49.6	99	30	143



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Surrogate Summary

SDG No.: O3631

Client: LaBella Associates P.C.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
O3631-10	RINSATE-BLANK	1,2-Dichloroethane-d4	50	50.6	101	74	125
		Dibromofluoromethane	50	55.3	111	75	124
		Toluene-d8	50	45.1	90	86	113
		4-Bromofluorobenzene	50	50.9	102	64	133
VN0714WBL01	VN0714WBL01	1,2-Dichloroethane-d4	50	49.2	98	74	125
		Dibromofluoromethane	50	56.7	113	75	124
		Toluene-d8	50	45.7	91	86	113
		4-Bromofluorobenzene	50	51.0	102	64	133
VN0714WBS01	VN0714WBS01	1,2-Dichloroethane-d4	50	52.0	104	74	125
		Dibromofluoromethane	50	51.5	103	75	124
		Toluene-d8	50	51.1	102	86	113
		4-Bromofluorobenzene	50	50.9	102	64	133
VN0714WBSD0	VN0714WBSD01	1,2-Dichloroethane-d4	50	53.6	107	74	125
		Dibromofluoromethane	50	54.0	108	75	124
		Toluene-d8	50	51.4	103	86	113
		4-Bromofluorobenzene	50	52.1	104	64	133